

Question: What does CLIA require regarding sinks, eye-wash stations, etc. in the lab?

Answer: Subpart J of 42 CFR 493 speak to facility requirements. This section does not state specific numbers of sinks or eye-wash stations as a requirement. Please read the section for details of the specific requirements and note that 42 CFR 492.1101 (c) does say that the laboratory must be in compliance with applicable Federal, State and local requirements. An example of requirements that must be met would include Federal and State OSHA requirements.